



Building the Road to Tomorrow: Implementing and Validating TB Laboratory-Developed Tests



Department of Health
Wadsworth Center

Building the Road to Tomorrow: Implementing and Validating TB Laboratory-Developed Tests

Kimberlee Musser, PhD

**APHL 14th National Conference on Laboratory Aspects of Tuberculosis
2026| March 10-12, 2026**



Some basics...

Laboratory Developed Test- Diagnostic test that is designed, manufactured, and performed entirely within a single, CLIA/CAP-certified laboratory.

Validation- Evidence-based process of ensuring a test, measurement, or system accurately fulfills its intended purpose, requirements, and user needs.

Implementation of an LDT- Differences between CLIA, CAP, NYS CLEP

- Clinical Director
- CAP Checklist
- CLEP Microbiology Molecular Checklist and Submission



Why do we need LDTs for TB Testing?

Very few FDA approved options for identification:

- MALDI Biotyper CA System
- Cepheid Xpert MTB/RIF Assay



Implementing and Validating TB Laboratory-Developed Tests



- FDA Modification
- Specimen type
- Specimen volume



- Validating all aspects of an NGS
- Designing a validation study and controls for a highly multiplexed PCR assay



Department of Health
Wadsworth Center

Clinical Validation and LDT Test Approval



Standard Operating Procedure

- Intended use
- Limitations
- Step-by-step procedures
- Reports
- Interpretation
- Quality control (QC)
- Assay controls
- Proficiency Testing

Assay Validation Study

- Specificity
- Sensitivity- Limit of Detection
- Reproducibility: Intra- & Inter-assay
- Accuracy
 - Retrospective studies
 - Prospective studies
 - Comparison to gold standard

Challenges/Questions

- How many samples do you test?
- Where can you get samples from?
- How do you handle discordance?
- What if there is no gold standard to compare to?
- What if validation is cost prohibitive?



Building the Road to Tomorrow: Implementing and Validating TB Laboratory-developed Tests

1:30 pm – 2:30 pm • 588-706-25 • contact hours: 1.0

Moderator: Kimberlee Musser, PhD, Chief of Bacterial Disease, New York State Department of Health-Wadsworth Center

Speakers:

- Yvette Vergnetti, MS, MLS-L, MT/MLS(ASCP), Mycobacteriology Technical Supervisor, Alaska Department of Health
- John Dulany, Senior Public Health Scientist, TB and Mycology, Arizona State Public Health Laboratory
- Matthew Sylvester, PhD, PMH-CA(AAB) Research Scientist Supervisor, Microbial Diseases Laboratory, California Department of Public Health



Beyond Sputum: Validating Additional Specimen Types for GeneXpert MTB/RIF



LDT Validation of Fluoroquinolones for MGIT & Agar Proportion DSTs



Laboratory Developed Test Validation: Targeted Next Generation Sequencing (tNGS)



Department of Health
Wadsworth Center



Beyond Sputum: Validating Additional Specimen Types for GeneXpert MTB/RIF

Yvette Vergnetti, MS MLSL, MT/MLS (ASCP)
TB Technical Supervisor

Alaska State Public Health Laboratories (ASPHL) - Anchorage


Alaska TB

Highest TB Rates in the nation

Reported Tuberculosis (TB) Cases


In 2024, 91 cases of TB disease were reported in Alaska, reflecting an incidence rate of 12.3 cases per 100,000 population. This represents an 18% increase in cases from the previous year. The U.S. TB incidence rate for 2024 was 3.0 cases per 100,000, with an 8% increase in cases from 2023.

Tuberculosis Incidence Rates, Alaska & U.S., 2014-2024



https://health.alaska.gov/media/mksh4drw/2024-ak-tb-summary-brief-final_111020255.pdf

Regional Incidence per 100,000 population



Alaska population 739,000

77% of Alaska's 2024 TB cases were among Alaska Native communities
16% Foreign-born

Alaska State Public Health TB Laboratory

- ▶ Only Mycobacteriology laboratory in Alaska
 - ▶ Statewide reference laboratory serving hospitals (urban, rural, military), public health clinics and the Health Department
 - ▶ CLIA-certified high-complexity laboratory with BSL3 containment
 - ▶ 3 FTEs supporting all TB diagnostics
 - ▶ Annual testing volume
 - ▶ >5,800 specimens processed annually (>7,000 in 2025)
 - ▶ >1,000 TB NAAT performed annually (both pulmonary and extrapulmonary)
 - ▶ For respiratory AFB smear negative – 2 TB NAATs are performed
 - ▶ Full-spectrum TB diagnostics: smear microscopy, solid & liquid culture, GeneXpert NAAT (TB ID), MALDI-TOF ID (MTB/NTM), and first-line drug susceptibility testing including FQs
 - ▶ Sequencing for drug resistance (including *pncA*) validation complete – launching next quarter

TB NAAT in Alaska


- ▶ LDT TB NAAT implemented in 2011
 - ▶ Validated for all sources except for blood, CSF, and stool
 - ▶ Labor intensive and time consuming (~4 hours hands-on per testing day)
 - ▶ 2-3 FTEs
 - ▶ Batched testing 3 times a week
- ▶ GeneXpert MTB/RIF assay 2016
 - ▶ FDA approved for sputum on patients ≥ 18 years of age only
 - ▶ More expensive than LDT
- ▶ Maintained both testing platforms to accommodate extrapulmonary and pediatric sputa
 - ▶ Not sustainable
 - ▶ Cartridge cost justified by improved workflow and freed staff time

GeneXpert MTB/RIF Validation - Pulmonary

- ▶ 2017 a pre-validation feasibility study performed on bronchial lavage/wash and tracheal aspirates
 - ▶ Demonstrated acceptable performance prior to formal LDT validation
- ▶ Validation Design
 - ▶ **Acceptance criteria:** Pulmonary specimens must meet package insert sensitivity for AFB smear and sputum collection method (<3,000 CFU)
 - ▶ 20 non-sputum (clinical & spiked)
 - ▶ Banked specimens stored at -80°C
 - ▶ Negative pooled processed sediments confirmed culture-negative
 - ▶ Spiking strategy/target
 - ▶ 0.5 McFarland suspensions of 1.5×10^6 CFU/mL of MTB RIF susceptible and resistant strains
 - ▶ Serial dilutions 10^{-1} , 10^{-2} , 10^{-3} , and 10^{-4}
 - ▶ Dilutions 10^{-2} through 10^{-4} tested

Limit of Detection (LoD)

Figure 2 – Addition of known MTC dilution to pulmonary specimen aliquot



For each dilution:

- ▶ 1 mL dilution combined with 1 mL pooled culture-negative sediment
- ▶ 1 AFB smear prepared, stained, and examined
- ▶ 2 GeneXpert MTB/RIF cartridges tested
- ▶ 2 Middlebrook 7H11 plates inoculated (100 μ L each) for colony enumeration
 - ▶ Observed CFUs significantly < 3,000

Results

- ▶ Analytical Sensitivity:
 - ▶ Quantitative culture demonstrated starting suspension aligned closer to $\sim 3.75 \times 10^6$ CFU/mL (vs. intended 1.5×10^6)
 - ▶ Consistent with McFarland variability of >0.5 (clumping, settling, manual turbidity reading)
 - ▶ Assay detection met package insert sensitivity target ($<3,000$ CFU)
- ▶ Accuracy: 100% agreement between MTB/RIF and ASPHL reference methods
- ▶ Precision: 9 specimens tested by two techs on different days; 100% reproducibility
- ▶ Analytical Specificity: NTM panel tested; no cross-reactivity observed
- ▶ Conclusion: Validation confirmed GeneXpert MTB/RIF as a reliable platform beyond sputum, enabling sustainable expansion of rapid TB diagnostics in Alaska

Lessons Learned from Pulmonary Validation

- ▶ Pilot first, validate second - small feasibility studies identify issues early
- ▶ Target concentrations are theoretical; plate counts are reality
 - ▶ Dilution math assumes ideal mixing - MTB clumping, suspension settling, and minor technique differences introduce variability which can shift CFU dramatically
- ▶ Standardize McFarland preparation and pipetting technique (be consistent)
 - ▶ For example – densitometer vs manual reading
- ▶ Decide and document variables in the validation plan
 - ▶ For example - If you don't you can end up changing prep methods mid-study
- ▶ Unexpected results are data - not failure
- ▶ Clear documentation supports regulatory compliance

Expansion of GeneXpert Validations

- ▶ 2018 MTB/RIF Validations
 - ▶ Pediatric sputum
 - ▶ Gastric aspirates
 - ▶ Tissue
 - ▶ Body fluids (excluding blood, CSF, urine, and stool)
- ▶ 2022 – MTB-positive culture confirmation from solid and liquid media
- ▶ What's next - Stool
 - ▶ Critical for pediatric and non-expectorating patients
 - ▶ Less invasive; easier collection in remote communities
 - ▶ Supported by WHO guidance
 - ▶ Without validation, patients lack access to this testing

Thank You

- ▶ Acknowledgements

- ▶ ASPHL Mycobacteriology Team
- ▶ ASPHL Laboratory Leadership
- ▶ Alaska TB Program
- ▶ Infectious Disease Specialists and Clinical Partners
- ▶ Statewide submitting facilities and clinicians

- ▶ Contact Information

Yvette Vergnetti

- ▶ Email: yvette.vergnetti@alaska.gov
- ▶ Phone: 907-334-2119



Laboratory Developed Test Validation: Targeted Next Generation Sequencing (tNGS)

Matthew Sylvester, PhD, PHM

Unit Supervisor

Mycobacterial Drug Susceptibility Testing Unit

Mycobacterial, Mycotic, & Parasitic Diseases Section

Microbial Diseases Laboratory Branch (MDL)

California Department of Public Health

matthew.sylvester@cdph.ca.gov


**14th National Conference on
Laboratory Aspects of
Tuberculosis**

March 11, 2026

Disclosures

- **Relevant Financial Relationships**
 - I have nothing to disclose
- **The findings and conclusions in this presentation are those of the author and do not necessarily represent the views or opinions of the California Department of Public Health (CDPH) or the California Health and Human Services Agency.**
- **The use of brand names and/or any mention or listing of specific commercial products or services herein is solely for educational purposes and does not imply endorsement by CDPH or our partners, nor discrimination against similar brands, products, or services not mentioned.**

Targeted NGS (tNGS) LDT Assay




tNGS at CDPH-MDL

Deplex Myc-TB
(GenoScreen, France)
(May, 2024)

Unavailability of Reagents:
Test Discontinued
(January, 2025)

Modified CDC
tNGS Assay
(March, 2026)

Overall TB-DST Workflow



Targeted NGS LDT Assay Design: Extraction and RT-PCR

Extraction



Add sample
to tube with
beads



Heat kill at 90°C for
30min



Bead Beat on
Fastprep



Purify on
Promega Maxwell

RT-PCR



Screening RT-PCR on BDMax Platform with BioGX reagents:

- Confirm presence of TB
- Confirm sufficient TB load for tNGS (Ct value)

Notes:

- Validated in parallel with tNGS
- Ct values may not reflect those of GeneXpert MTB/RIF

Targeted NGS LDT Assay Design: Illumina Library Prep and Sequencing

Singleplex PCRs (20) and
Illumina DNA Prep Pre-PCR



Post-PCR Clean-up



Pool and Load



iSeq 100 v3 kit

or




MiSeq i100
5M 300-cycle or
600-cycle kits

QC Checkpoint: after targeted
PCR pooling and clean-up


QC Checkpoint: After library
clean-up

Max: 6 samples + External Positive Control + External Negative Control

Bioinformatics Analysis



CLIA-Compliant LDT Validation Timeline



 Status: FINAL	California Department of Public Health Microbial Diseases Laboratory (MDL) Mycobacterial, Mycotic & Parasitic Diseases Section (MMPDS)	Validation Report
		MDL-MMPDS-DST-VR-003
		Page 1 of 129

+28 Appendices!!!

Validation of MTBC mCDC tNGS-DST - Report

[TITLE: **Validation of Targeted Next Generation Sequencing-Based
Drug Susceptibility Testing of *Mycobacterium Tuberculosis*
Complex from Sediments and Cultures Using Modified Centers
for Disease Control and Prevention Primer Set - Report**

Summary Performance Parameters

Performance Characteristic	Expected Values
Accuracy	≥90%
Diagnostic Sensitivity	≥90%
Diagnostic Specificity	≥90%
Analytical Sensitivity (Limit of detection--LOD)	N/A
Analytical Specificity	Cross-reactivity: None Interference: None
Reportable Genomic Range	Within primer boundaries
Precision (Repeatability & Reproducibility)	100%

Reference Methods

Validation Objective	Reference Method	Performance Parameters Calculated for:
Validity of tNGS Susceptible(S) / Resistance(R) predictions	Validated pDST, WGS S/R predictions, Deeplex Myc-TB tNGS S/R predictions	Each drug and overall
Validity of detected genomic variations (e.g. SNPs & InDels)	Previously-validated sequencing methods (WGS, Deeplex Myc-TB tNGS)	Each variation in the reportable genomic range (determined during validation)
Accurate identification of <i>M. tuberculosis</i> complex (not <i>M. bovis</i>) vs. <i>M. bovis</i> .	Validated sequencing methods (Pyrosequencing, WGS, Deeplex Myc-TB tNGS)	Each sample


Sample Types for Validation (Accuracy, Sensitivity, Specificity)

Type of validation material	Patient Material Source	MTBC status	N	Additional performance parameter	Covered Variability
Clinical Sediments: From CDPH, Kaiser Regional Laboratory	Pulmonary	MTBC-positive	34		1+ to 4+ AFB grades.
		MTBC-negative	2 (NTM neg)		
	Extrapulmonary samples (body fluid, processed tissue)	MTBC-positive	20		
		MTBC-negative	10 (5 NTM neg, 5 NTM pos)	Cross-reactivity (NTM)	
Contrived samples: Cultures spiked into negative sediments at ~1 log above LOD (1:1000 of a 0.5McF)	Pulmonary	MTBC-positive	18	Repeatability /Reproducibility	Increased variety: Mostly R; <i>M. bovis</i> and <i>M. bovis</i> BCG; Naturally occurring heteroresistance; Low-level RIF-R.
Cultures	N/A (extracted from pure culture)	MTBC strains	5 strains (14 tests with 3 strains x 4 media types)	MTBC ID panel: additional <i>M. bovis</i> & <i>M. bovis</i> BCG	Lowenstein-Jensen [LJ], 7H10, 7H11, early-positive MGIT
	N/A (simulated mixed cultures with NTM)	MTBC-positive	12	Cross-reactivity (NTM)	Mixed cultures

Samples (N=91 unique) were chosen to cover a range of representative mutations and drug resistance

Limit of Detection (LOD): Biological

- *In vitro* spike-in
 - H37Rv dilutions spiked into negative sediment
 - The lowest amount of MTBC yielding successful sequencing of all loci as well as accurate identification of genomic variations, correct S/R profile prediction, and MTBC ID in 95% of contrived specimens.
 - Additional contrived samples made at ~1 log above the LOD to assess other performance parameters
- *Clinical specimens* →
 - The lowest amount of MTBC (AFB gradin or NAAT Ct values) yielding successful sequencing in 95% of clinical specimens.
 - LOD determined to be Ct=32.5 from the BDMax RT-PCR assay (**Note: may not correlate with GeneXpert**) when combined with results from contrived samples



	All Ct values		Ct ≤Cutoff Threshold	
Specimen Category	Total	% Succeeded	Total	% Succeeded
Sputum	37	91.9%	31	100.0%
Sterile body fluid	9	100.0%	9	100.0%
Tissue	25	76.0%	16*	87.5%*
Urine	6	66.7%	4	100.0%
Total	77	85.7%	60	96.7%

*Matrix associated with sequencing failures but not reportable false variant calls

Limit of Detection (LOD): Bioinformatic

Variant Read Support: minimum read support permitting accurate identification of genomic variations in all tNGS DST loci with at least 95% accuracy.

Method:

- Clinical WGS sequences downsampled bioinformatically

Heteroresistance: Minimum allele frequency permitting accurate identification of genomic variations in all tNGS DST loci with at least 95% accuracy.

Method:

- Clinical samples with low-level heteroresistance
- Mixed WT and mutant DNA at AF=10%, 15%, 20%, 30%, undiluted

Parameter	Method	N datapoints	Obtained range	FP	FN	LoD95 (probit)
Read support	Bioinformatic downsampling	908	8x-165x	0	196	19.65 x
Allele frequency	Different ratios of WT and mutant DNA mixed <i>in vitro</i>	60	9.5% - 100%	0	8	11.2 % SNPs / 16.7% InDels

- 200x coverage permits detection of variants down to ~10% allele frequency
- 99.75% of reportable loci had coverage >200x

Cross-Reactivity with Nontuberculosis Mycobacteria

MTBC DNA	NTM DNA	NTM	rpoB	rrs	rrl
90%	10%	M. avium	p.Ser431Ile; p.Ser431Tyr; p.Leu443Ser; p.Leu443Phe; p.Asp190Glu; p.Ser239Asn; p.Val243Thr; p.Val243Ile; p.Arg253fs; p.Ser254fs; p.Val262Ala; p.432; p.Gln432Gln; p.Leu443Leu; p.Arg448Arg	n.1259C>T; n.1260G>A; n.1278A>T; n.1300C>T; n.1321G>A; n.1445C>T	No mutations detected
50%	50%	M. avium	p.Ser431Pro; p.Leu443Phe; p.Asp190Glu; p.Ser239Asn; p.Val243Thr; p.Val243Ala; p.Arg253fs; p.Ser254fs; p.Val262Ala; p.Val355Ile; p.432; p.Gln432Gln; p.Leu443Leu; p.Arg448Arg	n.1259C>T; n.1260G>A; n.1278A>T; n.1300C>T; n.1321G>A; n.1445C>T	No mutations detected
10%	90%	M. avium	p.Ser431Pro; p.Ser431Cys; p.Leu443Pro; p.Leu443Phe; p.Asp190Glu; p.Ser239Asn; p.Val243Thr; p.Val243Ile; p.Val243Ala; p.Arg253fs; p.Ser254fs; p.Val262Ala; p.Val355Ile; p.432; p.Gln432Gln; p.Arg448Arg	n.1259C>T; n.1260G>A; n.1278A>T; n.1300C>T; n.1321G>A; n.1445C>T	No mutations detected
90%	10%	Mix: M. abscessus, M. kansasii, M. fortuitum	p.Thr184Ser; p.Asp190Glu; p.Ser201Asn; p.Ser201Ala; p.Ser201Gly; p.Ser201Cys; p.Ser239Asn; p.Arg253fs	n.1278A>T	n.2331A>G
50%	50%	Mix: M. abscessus, M. kansasii, M. fortuitum	p.Thr184Ser; p.Asp190Glu; p.Ser201Asp; p.Ser201Gly; p.Ser201Cys; p.Ser239Asn; p.Arg253fs; p.Val262Ala; p.Gly426Gly; p.Phe433fs; p.Gln432fs; p.Leu443Leu	n.1278A>T	n.2002G>A; n.2331A>G
10%	90%	Mix: M. abscessus, M. kansasii, M. fortuitum	p.Thr184Ser; p.Asp190Glu; p.Ser201Gly; p.Ser201Cys; p.Ser239Asn; p.Val243Asn; p.Val243Thr; p.Val243Ile; p.Arg253fs; p.Val262Ala; p.Thr264Pro; p.Met387Leu; p.Gly426Gly; p.Phe433fs; p.Gln432fs; p.Gly442Gly; p.Leu443Leu	n.1278A>T	n.2002G>A; n.2331A>G; n.2567A>G; n.2594T>C; n.2624T>C; n.2640C>T; n.2642C>T; n.2872A>G; n.2923G>A; n.2940G>A; n.2962C>T; n.2971T>A; n.3022T>C

False positives detected for: *rpoB*, *rrs*, *rrl*, & *rplC*

- Often cluster
- Single false variants may be present in other genes

Thresholds set to minimize false variants:

- For mutations with allele frequency (AF) = 10%-94.99% => min 500x coverage required
- For mutations with allele frequency (AF) ≥ 95% => 100x coverage required


Rule for reporting: Suppress reporting for all of these loci if 4 or more reportable mutations appear in one locus => 100% of FP mutations filtered

Interfering Substances

Substance	Results
Whole blood	No inhibition observed up to 10% (v/v)
Specimen Processing Reagents (i.e. AlphaTec NAC-PAC EA3)	Loci failed sequencing (low coverage) with 125µl residual processing reagents
Rifampin	No inhibition observed up to 25 µg/ml
Ethambutol	False positives observed at 25 µg/ml

Reportable Genomic Range

- Criteria for inclusion of a position into the *broad* reportable genomic range:
- Associated with correct S/R profile in $\geq 90\%$ of samples and correct variant detection in 95% of samples
 - At least 90% of validation samples must have had the position covered by at least 50x



- Of n=839 locus-level observations:
- 98.33% of observations have BRR=100%
 - 99.76% have ERR=100%

Performance Parameter Results by Drug (Compared with Phenotypic DST)

Drug	N	TP	TN	FN	FP	Accuracy	Sensitivity	Specificity	PPV	NPV
INH	95	32	63	0	0	100%	100%	100%	100%	100%
ETO	77	25	52	0	0	100%	100%	100%	100%	100%
RIF	97	21	76	0	0	100%	100%	100%	100%	100%
PZA	97	31	65	1*	0	99.00%	96.90%	100%	100%	98.50%
EMB	99	17	82	0	0	100%	100%	100%	100%	100%
AMK	83	9	74	0	0	100%	100%	100%	100%	100%
KAN	81	9	72	0	0	100%	100%	100%	100%	100%
CAP	91	8	83	0	0	100%	100%	100%	100%	100%
MXF	65	14	51	0	0	100%	100%	100%	100%	100%
LFX	80	14	66	0	0	100%	100%	100%	100%	100%
BDQ	81	1	80	0	0	100%	100%	100%	100%	100%
CFZ	84	1	83	0	0	100%	100%	100%	100%	100%
LZD	73	0	73	0	0	100%	n/a	100%	n/a	100%
Total	1103	182	920	1	0	99.90%	99.50%	100%	100%	99.90%

*Mutation failed read support

- Also compared genomic variations

Summary Performance Parameters

Performance Parameter	Expected Values	Achieved Values
Accuracy	≥90%	99.9% (S/R profile); 99.94% (Genomic Variations); 98% (MTBC ID)
Diagnostic Sensitivity	≥90%	99.5% (S/R profile); 100% (Genomic Variations); 98% (MTBC ID)
Diagnostic Specificity	≥90%	100% (S/R profile); 99.94% (Genomic Variations); 100% (MTBC ID)
Analytical Sensitivity (Limit of detection)	N/A	AFB 2+ (pulmonary); AFB 3+ (tissues); Ct=32.5 based on 95% of samples sequencing successfully
Analytical Specificity	Cross-reactivity: None Interference: None	Cross-reactivity with NTM for <i>rpoB</i> , <i>rrs</i> , <i>rrl</i> , <i>rplC</i> Interference observed with a FP mutation with EMB
Reportable Genomic Range	Within primer boundaries	Broad reportable range and essential for reporting range defined for each locus
Precision (Repeatability & Reproducibility)	100%	100%

Acknowledgements

- **CDPH Infectious Diseases Laboratory Division:**
 - Zenda Berrada, MDL Lab Director
 - **Varvara Kozyreva, Mycobacterial, Mycotic, and Parasitic Diseases Section Chief**
 - **Tanya Holbrook**
 - Maftuna Gafurova
 - Josellin Figueroa-Sandoval
 - Cailin Mirabella
 - Alessandra Wetmore
 - Christian Lopez-Pena
 - Terry Weber
 - Angus Angermeyer
- **Kaiser Permanente**
 - Ali Aljanabi
- **Theiagen Genomics:**
 - Sage Wright
 - Theron James
 - Andrew Lang
 - Deborah Young, Krisandra Allen
- **CDC:**
 - Angela Starks
 - Jamie Posey
 - Lauren Cowan
- **APHL:**
 - Sarah Buss
- **ARUP**
 - Salika Shakir
- **Wadsworth Center:**
 - Kim Musser
 - Michelle Dickinson
 - Joseph Shea
 - Vincent Escuyer
- **Jody Phelan, London School of Hygiene and Tropical Medicine**
- **Florida Bureau of Public Health Laboratories**
 - Marie-Claire Rowlinson
 - Patrick Valois
 - Calin Chiribau
 - Matthew Schimenti

