

Traceability Quiz

Passing score = 80%

Select the correct answer.

1. T F ISO/IEC 17025 does not require laboratories to have calibration systems in place.
2. T F ISO Guide 34 specifies requirements for reference material producers.
3. T F Calibration laboratories do not need to establish traceability.
4. T F Calibration laboratories must issue certificates that include measurement uncertainty.
5. T F Traceability does not apply to instruments and equipment.
6. T F Traceability can be referred to as a “calibration hierarchy”.

7. T F Calibration is the comparison of two measurement devices or systems, one of known uncertainty (reference standard) and one of unknown uncertainty (test equipment).

8. Reference Standards include:

A. Hard and non-consumable items	B. Liquid standards	C. Consumables	D. Proficiency testing materials
----------------------------------	---------------------	----------------	----------------------------------

9. Reference Materials include:

A. Hard and non-consumable items	B. NIST traceable thermometers	C. Consumables	D. Weights
----------------------------------	--------------------------------	----------------	------------

10. Reference Standards should be:

A. Traced to SI units	B. Used for calibration only to maintain accuracy and integrity	C. Calibrated by an ISO/IEC 17025 accredited body that can provide traceability	D. B and C
-----------------------	---	---	------------

11. Reference Materials should be:

A. Traced to SI units	B. Used for calibration only to maintain accuracy and integrity	C. Calibrated by an ISO/IEC 17025 accredited body that can provide traceability	D. Made by the laboratory in-house for best results
-----------------------	---	---	---

12. Intermediate checks need defined procedures and schedules for carrying out checks to maintain confidence in the calibration status of:

A. Reference Standards and Materials	B. Primary and Working Standards	C. Transfer Standards	D. All of the choices
--------------------------------------	----------------------------------	-----------------------	-----------------------