



Public Health Service
DEPARTMENT OF HEALTH & HUMAN SERVICES

January 9, 2024

To: All recipients of the CDC Triplex Real Time RT-PCR

From: The CDC Dengue Branch, Triplex Team

Dear CDC Triplex Real Time RT-PCR user,

The CDC Triplex RT-PCR Assay, positive controls kit (KT-0167) Lot #: 19-0061 will be approaching its expiration date soon according to the product insert (09/04/2023) and most recent extension (01/04/2024). The CDC has performed QC evaluations of these kits (lot) and determined that their performance has not changed during the time they have been stored at -20 °C in our facilities. We have therefore extended the shelf life of these kits for another 6 months.

If you have kits from Lot #:19-0061, we ask that you ignore the expiration date on these kits and continue to use Lot #:19-0061 until 06/04/2024.

Sincerely,

A handwritten signature in black ink, appearing to read "Gilberto A. Santiago", is positioned below the word "Sincerely,".

Gilberto A. Santiago, Ph.D.
Lead, Molecular Diagnostics and Research
Surveillance and Research Laboratory
Division of Vector Borne Diseases
Centers for Disease Control and Prevention
1324 Calle Cañada
San Juan, PR 00920
TEL: (787) 706-2399
FAX: (787) 706-2496